

Equilibrium Conformations of Concentric-Tube Continuum Robots

D. Caleb Rucker, Robert J. Webster III,
Gregory S. Chirikjian, and Noah J. Cowan

Abstract

Robots consisting of several concentric, preshaped, elastic tubes can work dexterously in narrow, constrained, and/or winding spaces, as are commonly found in minimally invasive surgery. Previous models of these ‘active cannulas’ assume piecewise constant precurvature of component tubes and neglect torsion in curved sections of the device. In this paper we develop a new coordinate-free energy formulation that accounts for general preshaping of an arbitrary number of component tubes, and which explicitly includes both bending and torsion throughout the device. We show that previously reported models are special cases of our formulation, and then explore in detail the implications of torsional flexibility for the special case of two tubes. Experiments demonstrate that this framework is more descriptive of physical prototype behavior than previous models; it reduces model prediction error by 82% over the calibrated bending-only model, and 17% over the calibrated transmisional torsion model in a set of experiments.

1 Introduction

An active cannula is a continuum robot consisting of concentric elastic tubes, each of which has a preset curved shape (see Figure 1). This thin continuum robot design is mechanically

simple, and has the ability to reach dexterously into confined or winding environments (1). The tubes of the cannula form a single “backbone” that elastically changes shape as they translate and rotate axially. Transmitting moments within the backbone in this way provides an alternative to prior continuum robot designs that use support disks with tendon wires (2; 3), elastic sleeves with embedded tendons (4), flexible push rods (5), or pneumatic actuators (6; 7) to apply bending moments. Despite its mechanical simplicity, computing equilibrium conformations of an active cannula (i.e., the forward kinematics problem) remains a challenge.

The idea of deriving dexterity from counter-rotated, pre-curved concentric tubes was introduced relatively recently. Note that the kinematic conditioning of larger scale robot manipulators (with complex linkages, cabling, and actuation) will typically exceed that of active cannulas, so in the present context “dexterity” simply describes the affordance of multiple degrees-of-freedom at the tip given only basal actuation. Loser *et al.* (8) developed a steerable needle composed of two fully overlapping precurved cannulas whose bases rotate (but do not translate) relative to one another to change needle curvature. Daum (9) patented a deflectable needle assembly in which a curved “catheter” is deployed through a rigid outer cannula. A similar system is the “Curved Multi-Tube” (CMT) of Furusho *et al.*, where the tubes do not translate, but are constructed such that a portion of the inner tube lies beyond the end of the outer (10; 11). Initial CMT models required infinite flexural rigidity of every outer tube with respect to the collection of smaller tubes inside it, and infinite torsional rigidity of all tubes.

Recently, the more general active cannula mechanism—telescoping, concentric tubes that both translate and rotate with respect to one another—has emerged, along with beam mechanics models that account for flexural and (limited) torsional elasticity (12; 13; 14; 15; 1). Assuming piecewise-constant precurvature and torsional rigidity in curved sections, these models describe active cannula backbone shape by balancing moments between component tubes. Under these assumptions, the resulting active cannula shape is comprised of a finite




Figure 1: A prototype active cannula made of four superelastic Nitinol tubes and one central wire (with three tubes and the wire visible).

sequence of mutually tangent circular arcs.

This new actuation strategy appears to be well suited for applications at the “meso-scale” ($\approx 0.1\text{--}100$ mm) that require thin, dexterous manipulators, including minimally invasive surgical procedures. Specific applications for which active cannulas have been proposed include accessing the lung via the throat (14; 1), transgastric surgery (12), fetal procedures (10), steering needles embedded in tissue (8; 13), cardiac procedures (13), and transnasal skull base access (12). An overview of several specific ways active cannulas might be used in medicine is given in (16). It is also possible in principle to construct very small active cannulas which may be useful in cell manipulation (17; 18; 19) and other microsurgical applications. In some of the above applications (e.g. lung), active cannulas will be used in air-filled cavities, and thus the free-space cannula models developed in this paper will apply directly.

When the cannula is embedded in tissue (e.g. when used as a steerable needle), a free space kinematic model will need to be coupled to a tissue model to predict cannula shape, as has been done for straight needles (20; 21; 22). When used as steerable needle in the tissue embedded case, active cannulas will have both benefits and drawbacks when compared to tip-based steering techniques that utilize tissue interaction forces to steer (23; 24). Benefits

include the potential to manipulate tissue intentionally using cannula shaft shape change, and the ability to leave one soft tissue medium, traverse open space, and enter another. Potential drawbacks include increased complexity of control and the fact that cannula preshaping may impose restrictions on achievable shapes within tissue to which tip-steered needles are not subject. Independent of whether an active cannula is used in tissue or free space (e.g. in the lung), the first step in practical implementation is being able to predict cannula shape in free space as a function of axial rotations and translations of the component tubes, the “joint variables” of the robot.

Prior models neglect torsional effects in curved sections, but a recent model incorporating torsional effects in straight sections of the device (12; 1) demonstrates the critical role of torsion in describing categorical features of the energy landscape, as well as in accurately predicting tip position (14; 1). For the specific experimental setup reported in (1), this model predicts the location of the cannula end point with an average accuracy of 3.0 mm. However, the difference between predicted and experimental cannula tip positions is not uniform over the workspace, and is worst (8.76 mm) near the workspace boundary. Moreover, in this region, the model in (1) indicates that torsion in the straight section is at its highest. This suggests that, although neglected by this “lumped-parameter” model, there may be high torsional moments in the curved sections. Support for this was found in the first model that accounted for torsion in curved sections of a two-tube cannula with circular precurvatures, an initial version of which is available in (25) and an enhanced presentation of which can be found in Section 4. A subsequent presentation of the closed form two-tube solution appeared in (26). Derivation of the multi-tube case with variable curvature and torsion that we present in this paper has also been developed concurrently and independently and presented simultaneously in (27) and (26), where different analytical methods are applied to achieve the same final equations. Thus, it has come to be known that torsion in curved sections is a significant phenomenon that should be modeled. This motivates the generalized free space model we present in this paper, which accounts for both bending and torsion throughout a

multi-tube cannula, and enables use of cannulas incorporating variable precurvature.

Significant progress toward such a general model for wire-driven continuum robots with a single flexible backbone has been made in recent years. A general formulation for the energy stored in a deformed elastic backbone is given and minimized using variational calculus in (28), and used for wire driven continuum robots in (29) and (30). In this paper, we adopt a similar strategy to describe the equilibrium conformations of multiple precurved concentric tubes. Other recent related work has considered slender beam deformation from a group theoretic perspective (31). There, a model was derived that can predict beam deflection due to an applied wrench, which reduces in the 2D pure bending case to classical results. This work shows the feasibility of approaching beam mechanics problems via group theory.

In this paper, we present a coordinate free formulation for the energy stored within an active cannula. With arc length (rather than time) taken as the independent variable, the stored elastic energy in the backbone is analogous to the kinetic energy of a free rigid body. This formulation explicitly accounts for both bending and torsion throughout the cannula. It also accounts for general (non-circular) precurvatures of component tubes. We present several specific examples that illustrate how the Euler–Lagrange and Euler–Poincaré equations can be applied to derive a set of differential equations, a solution to which corresponds to the minimum energy conformation of the cannula.

Furthermore, we show that, with appropriate assumptions, our formulation includes prior models as special cases. We also present an analysis of the shape of curved, concentric tubes under both bending and torsion, deriving an analytical solution for the 2-tube case, and demonstrating that the resultant cannula shape will be non-circular. The experimental contribution of this paper is a demonstration that the new modeling framework can reduce model prediction error by 82% over the prior bending-only model, and 17% over the prior transmisional torsion model in a simple set of experiments with a prototype active cannula. These results were obtained with calibrated model parameters in all three cases.

2 Formulating the Energy Functional

We begin by formulating an energy functional which describes the total elastic energy due to bending and torsion stored in all tubes of a general n -tube cannula. We first provide a convention for describing the curves which define the shape of the individual tubes and the overall cannula shape. We then describe the total elastic (strain) energy stored in a single tube. The general energy functional can then be obtained by summing these individual tube energies.

2.1 Assumptions

The energy formulation in this section is performed under the standard assumptions of Kirchhoff rod theory, a special case of Cosserat rod theory (see (32) for an in-depth treatment of both). Kirchhoff theory assumes inextensibility and neglects transverse shear strain, which are generally regarded as good assumptions for long thin rods like the tubes that make up our active cannula. We also neglect gravitational effects in this analysis, because they have little effect at the scales and stiffnesses involved in our work. This can be seen from standard cantilever beam theory, which predicts a tip deflection caused by gravity of only 60 microns for a single straight, horizontally cantilevered tube 100mm long, with an OD of 1.6mm, an ID of 1.3mm, and a Youngs modulus of 50GPa. Note that this is a highly conservative calculation because (1) the actual cannula will consist of several concentric tubes and thus have a higher bending stiffness (2) it will not generally be straight (3) it will not generally be horizontally cantilevered, and (4) the elastic modulus may actually be up to 75GPa (the manufacturer, NDC, Inc., quotes a range of 41-75GPa). Thus, while it will certainly be useful in future work to develop models that take into account external loading (e.g. to enable force control of cannulas interacting with tissue), it does not appear to be necessary to consider gravitational loading in free space models such as the one we derive in this paper. Similarly, we neglect friction as has been done in all active cannula models to date.

Qualitatively the authors have observed some frictional hysteresis in prototypes with tightly packed tubes, but do not observe any discernible hysteresis in the prototype described in the experimental section of this paper. It is likely that frictional effects will be complex functions of a number of parameters including curvature functions, arc lengths, tolerances between tubes, surface smoothness, lubrication, etc., and a detailed study of all such effects is left to future work. Quantitatively, the suitableness of all the assumptions listed above can be tested by comparing model predictions to experimental tip positions. We provide such a comparison in the experimental section of this paper.

2.2 Parameterization of Curves and Notation

Suppose that each tube in isolation is described by an arc-length-parameterized space curve $\mathbf{t}_i(s)$ for $s \in [0, 1]$, with $\mathbf{t}_i(0) = \mathbf{0}$. Let us attach a reference frame at each value of arc length, s , with the local z axis pointing along the tangent to the curve, $\dot{\mathbf{t}}_i = d\mathbf{t}_i/ds$, and with the origin of the reference frame at $\mathbf{t}_i(s)$. Furthermore, let us establish the local x and y axes in any canonical way. For example, we can use the Frenet–Serret apparatus, or use Bishop’s frames (33), etc., as reviewed in (34). In any case, this will mean that a set of reference frames $g_i(s) \in \text{SE}(3)$ will be defined, one for each tube as

$$g_i(s) = \begin{bmatrix} R_i(s) & \mathbf{t}_i(s) \\ \mathbf{0}^T & 1 \end{bmatrix}, \quad (1)$$

where $R_i(s) \in \text{SO}(3)$ is the rotation of the frame at the point s along the curve relative to the base frame at $s = 0$ and

$$\mathbf{t}_i(s) = \int_0^s R_i(\sigma) \mathbf{e}_3 d\sigma \quad (2)$$

is the translation of the frame.

Let us assume that the only degrees of freedom required to describe the conformations of

each tube under all possible deformations due to moderate forces are bending and twisting. In other words, if $R_i^*(s)$ defines the pre-curved shape then the only admissible deformations will be of the form $R_i^*(s) \rightarrow R_i(s) \in \text{SO}(3)$. In this model, arc length is not changed by deformation (implying that each cannula tube behaves as an inextensible rod), and the relationship between position and orientation in (2) applies to the deformed parameters $\mathbf{t}_i(s), R_i(s)$ as well as the original parameters $\mathbf{t}_i^*(s), R_i^*(s)$ (implying that there is no shear strain in the deformations).

In general, $R(s)^T \dot{R}(s) \in \mathfrak{so}(3)$ is a 3×3 skew symmetric matrix, that can be parameterized by a three-vector. The isomorphism $\mathbb{R}^3 \simeq \mathfrak{so}(3)$ is defined by

$$\hat{\wedge}: \begin{bmatrix} \omega_1 \\ \omega_2 \\ \omega_3 \end{bmatrix} \mapsto \begin{bmatrix} 0 & -\omega_3 & \omega_2 \\ \omega_3 & 0 & -\omega_1 \\ -\omega_2 & \omega_1 & 0 \end{bmatrix}, \quad (3)$$

where $\mathfrak{so}(3)$ is the Lie algebra of $\text{SO}(3)$. For $\Omega = \hat{\omega} \in \mathfrak{so}(3)$, the inverse is given by $\wedge: \Omega \mapsto \begin{bmatrix} \omega_1 & \omega_2 & \omega_3 \end{bmatrix}^T$. The unit vectors $\mathbf{e}_1, \mathbf{e}_2, \mathbf{e}_3 \in \mathbb{R}^3$ are the standard basis.

2.3 The Kinetic Analogy

Throughout this paper we will employ the well-known kinetic analogy from the mechanics literature, which enables one to discuss concepts from mechanics using language that is perhaps more familiar in the robotics community. For a review of the kinetic analogy see (35). In the context of this paper, use of the kinetic analogy has implications for interpretation of $\omega(s)$, which denotes a vector of arc-length-parametrized curvatures. For example, using any frame convention with its z axis tangent to the curve (which we do exclusively in this paper¹), $\omega(s)$ is a vector of three elements, the first two of which denote curvatures about the x and y axes, and the last of which denotes the torsion about the z axis. Equivalently, from the perspective of the kinetic analogy, $\omega(s)$ can be viewed as the angular velocity (parametrized

¹Although we note that our energy formulation is general enough to account for any frame choice.

by arc length rather than time) of a coordinate frame that travels along the centerline curve of the cannula. This distinction has no bearing on the mathematical equations presented in the remainder of this paper—the equations can be understood equally well from both perspectives.

2.4 Elastic Energy Stored in a Collection of Tubes

The deformation energy for an inextensible and shearless linearly elastic rod is given by

$$E = \frac{1}{2} \int_0^L [\boldsymbol{\omega}(s) - \boldsymbol{\omega}^*(s)]^T K(s) [\boldsymbol{\omega}(s) - \boldsymbol{\omega}^*(s)] ds, \quad (4)$$

where $\boldsymbol{\omega}^*(s) = (R^*(s)^T \dot{R}^*(s))^\vee$ is the local curvature (in the kinematic analogy the body “angular velocity” of the pre-shaped, unloaded tube frame $g(s)$, with arc length as the independent variable rather than time) and $\boldsymbol{\omega}(s) = (R(s)^T \dot{R}(s))^\vee$ is the local curvature after deformation. Equation 4 is an expression for the strain energy due to bending and torsion in a rod under the assumption of the Bernoulli–Euler constitutive law in which the internal moment is proportional to the change in curvature. It is also invariant to frame assignment, meaning that use of Frenet–Serret frames, Bishop’s frames, or another frame convention will yield the same total energy. A very similar formulation of elastic energy in a Kirchhoff rod can be found in (36), and a more general formulation (for a tendon-driven continuum robot) which includes extensibility and shear is available in (30). In general, the symmetric stiffness matrix in (4) depends on the material properties of the tube, its geometry, and the way in which the reference frame is attached. For example, if the local x and y axes of the reference frame are attached according to the Frenet–Serret apparatus, and in this system $K(s)$ is diagonal (and possibly also independent of s), then in another reference system, such as a Bishop’s frame which evolves along the backbone with minimal total orientational change, then $\tilde{K}(s) = Q^T(s)K(s)Q(s)$ where $Q(s) = \exp\{\hat{\mathbf{e}}_3\theta(s)\} \in \text{SO}(3)$ is the relative rotation between these two systems of backbone reference frames around the z axis. However, in the

special case that

$$K = \begin{bmatrix} k_1 & 0 & 0 \\ 0 & k_1 & 0 \\ 0 & 0 & k_3 \end{bmatrix} \quad (5)$$

is constant in one such coordinate system (i.e. when the tube has an annular cross section), then $\tilde{K}(s) = K$. Thus the stiffness matrix will be independent of the local orientation about the z axis of the coordinate systems, and thus the energy will be also. For a cylindrical tube of constant cross-sectional inertia, I , and polar moment, J , then $k_1 = EI$ and $k_3 = GJ$, where E denotes the Young's Modulus and G denotes the sheer modulus. Note that K can never be a scaled identity for cylindrical tubes, since

$$EI = GJ \Rightarrow \frac{E}{G} = \frac{J}{I} = 2 \Rightarrow 2(1 + \nu) = 2 \Rightarrow \nu = 0,$$

and, for physical materials, $\nu \neq 0$.

For a general n -tube cannula, we assume that all tubes overlap continuously for $s \in [0, L]$. Then, the stored elastic energy in entire device will be the sum of the energies stored in the individual tubes:

$$E = \frac{1}{2} \sum_{i=1}^n \int_0^L [\boldsymbol{\omega}_i(s) - \boldsymbol{\omega}_i^*(s)]^T K_i(s) [\boldsymbol{\omega}_i(s) - \boldsymbol{\omega}_i^*(s)] ds. \quad (6)$$

where $\boldsymbol{\omega}_i^*(s) = (R_i^*(s)^T \dot{R}_i^*(s))^\vee$ is the pre-shaped curvature and $\boldsymbol{\omega}_i(s) = (R_i(s)^T \dot{R}_i(s))^\vee$ is the equilibrium curvature for tubes $i = 1, \dots, n$. The tubes are concentric (and thus constrained to follow a common trajectory in space), implying $\mathbf{t}_i(s) = \mathbf{t}_1(s)$. Hence, there are only $2 + n$ independent degrees of freedom to account for in the final deformed cannula. Specifically, suppose the first tube is deformed via $R_1^*(s) \rightarrow R_1(s)$ (or equivalently $\boldsymbol{\omega}_1^*(s) \rightarrow \boldsymbol{\omega}_1(s)$, since $R_i^*(0) = R_i(0) = I$). This requires three degrees of freedom since $\boldsymbol{\omega}_1 \in \mathbb{R}^3$. As can be seen




Figure 2: Shown here are coordinate frames for the first and the i^{th} tubes at an arbitrary cross section of the active cannula. They differ by an angular rotation of θ about their Z axes, which are both tangent to the curve. The curvatures, ω and ω_i , of the frames are not pictured, since they may in general lie in any direction with respect to the coordinate frames.

from (2), all subsequent tubes must share the same z -axis, namely $R_i(s)\mathbf{e}_3 = R_1(s)\mathbf{e}_3$, to ensure that $\mathbf{t}_i(s) = \mathbf{t}_1(s)$ for $i = 2, \dots, n$. This leaves only one extra degree of freedom per tube: a rotation around the z -axis. Thus, the final conformation of each tube can be parameterized with a rotation $\theta_i \in \mathbb{S}^1 = \mathbb{R} \bmod 2\pi$, namely $R_i(s) = R_1(s) \exp \{\hat{\mathbf{e}}_3 \theta_i(s)\}$, $s \in [0, L]$, $i = 2, \dots, n$. Figure 2 below illustrates how $\theta_i(s)$ relates $g_1(s)$ to $g_i(s)$.

The curvatures of coordinate frames along these subsequent tubes will be related to that of the first tube via

$$\omega_i(s) = e^{-\hat{\mathbf{e}}_3 \theta_i(s)} \omega_1(s) + \dot{\theta}_i(s) \mathbf{e}_3, \quad i = 2, \dots, n. \quad (7)$$

Note that the energy functional (6) is invariant under actions of $\text{SO}(3)$: the rotation R_1 does not appear in the expression. The objective of this paper is to minimize (6) over all paths

$$(\omega_1(s), \theta_2(s), \dot{\theta}_2(s), \dots, \theta_n(s), \dot{\theta}_n(s)) \in \mathbb{R}^3 \times T\mathbb{T}^{n-1} \quad (8)$$

on the interval $s \in [0, L]$, where $\mathbb{T}^k = \mathbb{S}^1 \times \dots \times \mathbb{S}^1$ (k times) denotes a k -torus, and $T\mathbb{T}^k \simeq \mathbb{T}^k \times \mathbb{R}^k$ denotes its tangent bundle. Here, \mathbb{R}^3 is identified with the tangent space of

$\text{SO}(3)$ via left translation, i.e. $T\text{SO}(3) \ni (R, \dot{R}) \mapsto (R, R^T \dot{R}) \mapsto (R, \boldsymbol{\omega}) \in \text{SO}(3) \times \mathbb{R}^3$.

3 Two Special Cases

We now examine two special cases of the minimization of (6) using a variational approach, and show that for these cases, the variational approach reduces to “point-wise” minimization of the integrand in (6). We begin with a simple illustrative example of two tubes with coplanar pre-curvature. The result is the same as that obtained for circular curves in (1), but here we generalize to variable planar curvature. We then consider the case where the two planar curves are rotated axially with respect to one another, but are infinitely torsionally rigid. Again, the result presented here for general curvatures agrees with previous results for piecewise constant curvatures, assuming no torsion (12; 13; 1). It is also possible to include torsion in the straight transmission alone for piecewise constant curvature (1) in a straightforward manner. Therefore we generalize in Section 4 and present the solution to the case in which two arbitrarily pre-shaped tubes are rotated axially at their bases with respect to one another, and undergo torsional deformation along the length of the curves. Unlike the special cases discussed in this section, solution to the general case requires a variational approach.

3.1 Planar Theory: A Degenerate Case

In the planar case, we can assume without loss of generality that all curvatures are of the form $\boldsymbol{\omega}(s) = \kappa(s)\mathbf{e}_1$ where $\kappa(s)$ is the curvature. The energy in this case will be

$$E_1 = \frac{1}{2} \int_0^L \left\{ k_1(\kappa(s) - \kappa_1^*(s))^2 + k_2(\kappa(s) - \kappa_2^*(s))^2 \right\} ds \quad (9)$$

where $k_i = E_i I_i$. If we minimize this integrand point wise with respect to $\kappa(s)$, the result is

$$\kappa(s) = \frac{k_1 \kappa_1^*(s) + k_2 \kappa_2^*(s)}{k_1 + k_2}. \quad (10)$$

This result is true independent of whether or not the stiffnesses depend on s , and it is also true independent of whether planar tube precurvatures are circular.

Note also that the same result can be obtained via variational methods. To see how this is true, simply replace κ with the equivalent $\dot{\theta}$, where θ is the counterclockwise-measured angle that the tangent to a curve makes with respect to a fixed line, and apply the Euler-Lagrange equation. In this case, the Euler-Lagrange equation reduces to $\partial f / \partial \dot{\theta} = c$ where c is an arbitrary constant. This gives a result identical to (10) up to the constant of integration c . One can then determine c by substituting the result back into the original energy functional and noting that if there are no constraints on $\dot{\theta}$, then E is minimized when $c = 0$. Thus, the pointwise and variational minimization approaches produce the same result.

3.2 The 3D Case With No Torsion

Next, consider an active cannula with two tubes that are infinitely torsionally rigid. That is, tube cross sections are not able to rotate relative to one another about their common z axes along their lengths, namely $\theta_2 \equiv 0$. This was the assumption in (1; 13) for sections of the active cannula in which $\omega_i^*(s) \neq 0$ for any of the overlapping tubes. With a proper choice of coordinate frames, this reduces the variables over which we are minimizing (8) to $\omega = \omega_1 = \omega_2$.

We seek the rotation matrix $R(s)$ (or equivalently the curvature $\omega = (R^T \dot{R})^\vee$ since $R(0) = I$ is known), and from this the corresponding curve defined by (2), such that the following energy is minimized:

$$E = \frac{1}{2} \sum_{i=1}^2 \int_0^L [\omega(s) - \omega_i^*(s)]^T K_i [\omega(s) - \omega_i^*(s)] ds. \quad (11)$$

Completing the square inside the integrand, we have

$$\begin{aligned} E &= \frac{1}{2} \int_0^L (\boldsymbol{\omega}(s) - \boldsymbol{\alpha}(s))^T K (\boldsymbol{\omega}(s) - \boldsymbol{\alpha}(s)) + C(s) ds \\ &= \frac{1}{2} \int_0^L (\boldsymbol{\omega}(s) - \boldsymbol{\alpha}(s))^T K (\boldsymbol{\omega}(s) - \boldsymbol{\alpha}(s)) ds + \text{const.}, \end{aligned}$$

where,

$$\begin{aligned} K &= K_1 + K_2, \\ \boldsymbol{\alpha}(s) &= K^{-1} (K_1 \boldsymbol{\omega}_1^*(s) + K_2 \boldsymbol{\omega}_2^*(s)), \quad \text{and} \\ C(s) &= \boldsymbol{\omega}_1^{*T} K_1 \boldsymbol{\omega}_1^* + \boldsymbol{\omega}_2^{*T} K_2 \boldsymbol{\omega}_2^* - \boldsymbol{\alpha}^T K \boldsymbol{\alpha}. \end{aligned}$$

Here, $C(s)$ is independent of $\boldsymbol{\omega}$, and its definite integral is a strictly positive constant. This can be seen by noting that the integrand in (11) is always nonnegative, so when $\boldsymbol{\omega} = \boldsymbol{\alpha}$, the only piece of the integrand remaining is $C(s)$, which must be greater than or equal to zero. Thus, $\boldsymbol{\omega}(s) = \boldsymbol{\alpha}(s)$ provides the global energy minimum by minimizing the integrand in a “point-wise” fashion. This is the curvature that minimizes the energy, and is identical to the solution determined in (12; 1) and (13) for piecewise constant curvature tubes. One could also employ a variational approach to derive the energy minimum, as we will in the general case of the next section, namely the Euler-Poincaré equations.

4 The Many-Tube Case with Torsion

We now turn our attention to the general problem of an active cannula with any number of tubes that are flexible in both bending and torsion. Here torsional deformation is allowed in the curved sections, unlike in prior work.

The stored elastic energy is given by (6). By using (7), the variables over which we are minimizing reduce to the set given in (8). To simplify notation, let $\boldsymbol{\omega}(s) = \boldsymbol{\omega}_1(s)$ denote the deformed curvature of the first tube and let $\mathbf{t}(s) = \mathbf{t}_1(s)$. Denoting $R_\theta = e^{\hat{\mathbf{e}}_3 \theta(s)}$, then from

(7) we have $\boldsymbol{\omega}_i = R_{\theta_i}^T \boldsymbol{\omega} + \dot{\theta}_i \mathbf{e}_3$.

Substituting (7) into (6), we have,

$$\begin{aligned}
E &= \frac{1}{2} \int_0^L [\boldsymbol{\omega} - \boldsymbol{\omega}_1^*]^T K_1 [\boldsymbol{\omega} - \boldsymbol{\omega}_1^*] + \sum_{i=2}^n [R_{\theta_i}^T \boldsymbol{\omega} + \dot{\theta}_i \mathbf{e}_3 - \boldsymbol{\omega}_i^*]^T K_i [R_{\theta_i}^T \boldsymbol{\omega} + \dot{\theta}_i \mathbf{e}_3 - \boldsymbol{\omega}_i^*] ds \\
&= \frac{1}{2} \int_0^L \boldsymbol{\omega}^T K_1 \boldsymbol{\omega} - 2\boldsymbol{\omega}^T K_1 \boldsymbol{\omega}_1^* + \boldsymbol{\omega}_1^{*T} K_1 \boldsymbol{\omega}_1^* \\
&\quad + \sum_{i=2}^n \left(\boldsymbol{\omega}^T K_i \boldsymbol{\omega} + 2\boldsymbol{\omega}^T K_i \dot{\theta}_i \mathbf{e}_3 - 2\boldsymbol{\omega}^T R_{\theta_i} K_i \boldsymbol{\omega}_i^* \right. \\
&\quad \left. + \boldsymbol{\omega}_i^{*T} K_i \boldsymbol{\omega}_i^* - 2\boldsymbol{\omega}_i^{*T} K_i \dot{\theta}_i \mathbf{e}_3 + \dot{\theta}_i^2 \mathbf{e}_3^T K_i \mathbf{e}_3 \right) ds. \tag{12}
\end{aligned}$$

Since we are concerned with the case of tubes with annular cross sections, K_i is diagonal with its first two elements equal, and so

$$\boldsymbol{\omega}^T R_{\theta_i} K_i R_{\theta_i}^T \boldsymbol{\omega} = \boldsymbol{\omega}^T K_i \boldsymbol{\omega}, \text{ and } \boldsymbol{\omega}^T R_{\theta_i} K_i \dot{\theta}_i \mathbf{e}_3 = \boldsymbol{\omega}^T K_i \dot{\theta}_i \mathbf{e}_3. \tag{13}$$

To simplify the above expression, let

$$\begin{aligned}
\bar{\boldsymbol{\omega}}_i^* &= R_{\theta_i} \boldsymbol{\omega}_i^* - \dot{\theta}_i \mathbf{e}_3, \\
K &= \sum_{i=1}^n K_i, \\
\boldsymbol{\alpha} &= K^{-1} (K_1 \boldsymbol{\omega}_1^* + \sum_{i=2}^n K_i \bar{\boldsymbol{\omega}}_i^*), \\
C &= \boldsymbol{\omega}_1^{*T} K_1 \boldsymbol{\omega}_1^* + \sum_{i=2}^n \bar{\boldsymbol{\omega}}_i^{*T} K_i \bar{\boldsymbol{\omega}}_i^* - \boldsymbol{\alpha}^T K \boldsymbol{\alpha}, \tag{14}
\end{aligned}$$

Here, $C = C(\theta_2, \theta_3, \dots, \theta_n, \dot{\theta}_2, \dot{\theta}_3, \dots, \dot{\theta}_n)$ is a leftover non-negative constant as before, but now it depends on the state variables, as does $\boldsymbol{\alpha} = \boldsymbol{\alpha}(\theta_2, \theta_3, \dots, \theta_n, \dot{\theta}_2, \dot{\theta}_3, \dots, \dot{\theta}_n)$. As with the no torsion case of Section 3.2, we can complete the square:

$$E = \frac{1}{2} \int_0^L (\boldsymbol{\omega} - \boldsymbol{\alpha})^T K (\boldsymbol{\omega} - \boldsymbol{\alpha}) + C ds. \tag{15}$$

In our approach, one applies the Euler–Poincaré equations on the integrand of (15) with respect to the Lie group $\text{SO}(3) \times \mathbb{T}^n$. For a derivation and review of the Euler–Poincaré equations see (37) and (38). This gives a system of differential equations in the variables $\boldsymbol{\omega}, \boldsymbol{\theta}, \dot{\boldsymbol{\theta}}$:

$$K(\dot{\boldsymbol{\omega}} - \dot{\boldsymbol{\alpha}}) + \boldsymbol{\omega} \times (K(\boldsymbol{\omega} - \boldsymbol{\alpha})) = 0 \quad (16)$$

$$\frac{\partial f}{\partial \theta_i} - \frac{d}{ds} \left(\frac{\partial f}{\partial \dot{\theta}_i} \right) = 0. \quad (17)$$

For illustrative purposes, we note that that (15) can be minimized point-wise by setting $\boldsymbol{\omega} = \boldsymbol{\alpha}$; this happens to be the trivial solution to (16).

Also, the application of static equilibrium conditions gives the same result (25). Writing a moment balance at an arbitrary cross section of the cannula and applying (7) and the definition of $\boldsymbol{\alpha}$, we have

$$K_1(\boldsymbol{\omega} - \boldsymbol{\omega}_1^*) + \sum_{i=2}^n R_{\theta_i} K_i(\boldsymbol{\omega}_i - \boldsymbol{\omega}_i^*) = 0 \implies \boldsymbol{\omega} = \boldsymbol{\alpha}. \quad (18)$$

The algebraic relation $\boldsymbol{\omega} = \boldsymbol{\alpha}$ provides $\boldsymbol{\omega}$ in terms of $\theta_2, \theta_3, \dots, \theta_n$ and $\dot{\theta}_2, \dot{\theta}_3, \dots, \dot{\theta}_n$, so we apply (17) (which is also the classical Euler–Lagrange equation, to the functional in (15) $n - 1$ times, once with respect to each θ_i .

Below, we apply Euler–Lagrange before substituting in the relationship $\boldsymbol{\omega} = \boldsymbol{\alpha}$. Substituting before applying Euler–Lagrange is equivalent to assuming that $\boldsymbol{\omega} = \boldsymbol{\alpha}$ is a holonomic constraint (39). This constraint turns out to be holonomic, which is verified by the fact that our result in the next section can be obtained by substituting before or after applying Euler–Lagrange. To apply the Euler–Lagrange equation to our functional f , we begin by first expanding f to obtain,

$$f = \boldsymbol{\omega}^T K \boldsymbol{\omega} - 2\boldsymbol{\omega}^T K \boldsymbol{\alpha} + \boldsymbol{\omega}_1^{*T} K_1 \boldsymbol{\omega}_1^* + \sum_{i=1}^n \bar{\boldsymbol{\omega}}_i^{*T} K_i \bar{\boldsymbol{\omega}}_i^*. \quad (19)$$

Based on this, the terms we need for the Euler-Lagrange equation are

$$\begin{aligned}\frac{\partial f}{\partial \theta_i} &= -2\boldsymbol{\omega}^T K \frac{\partial \boldsymbol{\alpha}}{\partial \theta_i} + 2\bar{\boldsymbol{\omega}}_i^{*T} K_i \frac{\partial \bar{\boldsymbol{\omega}}_i^*}{\partial \theta_i}, \quad \text{and} \\ \frac{\partial f}{\partial \dot{\theta}_i} &= -2\boldsymbol{\omega}^T K \frac{\partial \boldsymbol{\alpha}}{\partial \dot{\theta}_i} + 2\bar{\boldsymbol{\omega}}_i^{*T} K_i \frac{\partial \bar{\boldsymbol{\omega}}_i^*}{\partial \dot{\theta}_i}.\end{aligned}\tag{20}$$

Then, noting that

$$\frac{\partial \boldsymbol{\alpha}}{\partial \theta_i} = K^{-1} K_i \frac{\partial \bar{\boldsymbol{\omega}}_i}{\partial \theta_i}, \quad \frac{\partial \boldsymbol{\alpha}}{\partial \dot{\theta}_i} = K^{-1} K_i \frac{\partial \bar{\boldsymbol{\omega}}_i}{\partial \dot{\theta}_i}, \quad \text{and} \quad \frac{d}{dt} \left(\frac{\partial \bar{\boldsymbol{\omega}}_i}{\partial \dot{\theta}_i} \right) = 0,\tag{21}$$

we can simplify the Euler-Lagrange equation to

$$(\boldsymbol{\omega} - \bar{\boldsymbol{\omega}}_i^*)^T K_i \frac{\partial \bar{\boldsymbol{\omega}}_i^*}{\partial \theta_i} - (\dot{\boldsymbol{\omega}} - \dot{\bar{\boldsymbol{\omega}}}_i^*)^T K_i \frac{\partial \bar{\boldsymbol{\omega}}_i^*}{\partial \dot{\theta}_i} = 0\tag{22}$$

for $i = 2, \dots, n$. Then applying $\boldsymbol{\omega} = \boldsymbol{\alpha}$, we have a set of $n - 1$ differential equations describing $\theta_2, \dots, \theta_n$,

$$(\boldsymbol{\alpha} - \bar{\boldsymbol{\omega}}_i^*)^T K_i \frac{\partial \bar{\boldsymbol{\omega}}_i^*}{\partial \theta_i} - (\dot{\boldsymbol{\alpha}} - \dot{\bar{\boldsymbol{\omega}}}_i^*)^T K_i \frac{\partial \bar{\boldsymbol{\omega}}_i^*}{\partial \dot{\theta}_i} = 0.\tag{23}$$

where the boundary conditions are $\theta_i(0) = \theta_{i0}$ and $\dot{\theta}_i(\ell) = 0$ for $i \in \{2, \dots, n\}$.

We note that using the Euler-Poincaré equations provided a simple one-step way to obtain the governing equations, whereas the method of a static equilibrium balance combined with the Euler-Lagrange equation provided the same result via a more circuitous route. This becomes important as the complexity of the model increases. Since the Euler-Poincaré equations are simpler to apply and require less physical intuition, they may be particularly advantageous in future work which may include, for example, the effects of shear and extension.

4.1 Model Evaluation

Numerical evaluation of active cannula models has been previously addressed in (27) and (40). For completeness, we provide a brief review of these results here. In general, one can solve this system numerically for $\theta_2, \dots, \theta_n$. The expression (23) can be expanded in terms of precurvatures and stiffness matrices as follows,

$$\mathbf{e}_3^T \left(\sum_{j=2}^n \ddot{\theta}_j K_j K^{-1} - \ddot{\theta}_i \right) K_i \mathbf{e}_3 = \boldsymbol{\alpha}^T K_i \frac{\partial R_{\theta_i}}{\partial \theta_i} \boldsymbol{\omega}_i^* \quad (24)$$

In the case of $n = 2$ and constant curvature ($\boldsymbol{\omega}_i^* = \kappa_{xi} \mathbf{e}_1$) this expression reduces to the differential equation derived in (25). Equation (24) can be reformulated into a format amenable to numerical integration by first writing it in matrix form as,

$$\ddot{\boldsymbol{\theta}} = \mathbf{T}^{-1} \mathbf{f}(\theta_2, \dots, \theta_n), \quad (25)$$

Where $\mathbf{T} = \begin{bmatrix} t_{i,j} \end{bmatrix}$ is an $n-1 \times n-1$ symmetric torsional stiffness matrix given by

$$t_{i,j} = \begin{cases} \frac{-J_{i+1}G_{i+1} \sum_{k=1, k \neq i+1}^n J_k G_k}{\sum_{k=1}^n J_k G_k} & i = j \\ \frac{J_{i+1}G_{i+1}J_{j+1}G_{j+1}}{\sum_{k=1}^n J_k G_k} & i \neq j \end{cases} \quad (26)$$

Here,

$$\ddot{\boldsymbol{\theta}} = \begin{bmatrix} \ddot{\theta}_2 & \dots & \ddot{\theta}_n \end{bmatrix}^T, \text{ and} \\ \mathbf{f}(\theta_2, \dots, \theta_n) = \begin{bmatrix} f_1 & \dots & f_{n-1} \end{bmatrix}^T, \quad (27)$$

where,

$$f_i = \boldsymbol{\alpha}^T K_i \frac{\partial R_{\theta_i}}{\partial \theta_i} \boldsymbol{\omega}_i^*. \quad (28)$$

One can use the standard state definition procedure to write (25) in the form of a first order system $\dot{\mathbf{x}} = f(\mathbf{x})$ by augmenting the vectors on each side of the equation. Let $x_i = \theta_{i+1}$ for $1 \leq i \leq n - 1$ and let $x_i = \dot{\theta}_{i-n+2}$ for $n \leq i \leq 2n - 2$. The augmented first order system is then

$$\dot{\mathbf{x}} = \begin{bmatrix} \left[x_n, x_{n+1}, \dots, x_{2n-2} \right]^T \\ \hline \mathbf{T}^{-1} \mathbf{f}(x_1, x_2, \dots, x_{n-1}) \end{bmatrix}. \quad (29)$$

This system may now be solved with any of a number of numerical techniques for solving boundary value problems with mixed boundary conditions, for example Matlab's `bvp4c` function, or a shooting method which iteratively converges on the proper initial conditions to satisfy the boundary conditions. Since all the tubes are the same length here, the boundary conditions will be $x_i(0) = \theta_{i+1}(0)$ for $1 \leq i \leq n - 1$ and $x_i(\ell) = \dot{\theta}_{i-n+2}(\ell) = 0$ for $n \leq i \leq 2n - 2$.

Equation 29 governs the behavior of a single section of overlapping tubes. It is straightforward to apply this to an active cannula with multiple regions of overlap, where tubes end at different arc lengths. One can consider each unique region of tube overlap (between arc lengths where tubes end) separately. In each region, the variables obey (29) for the particular combination of tubes in that section. All that must be done to join these separate systems together is to enforce continuity of the boundary conditions at the junctions between systems and shift each distal end boundary condition to the end of its particular tube. This procedure is mathematically equivalent to defining a new system similar to (29), but which is piecewise defined by a system for each unique region of tube overlap (based on the presence and the curvedness of the tubes). However, this piecewise defined differential equation will be discontinuous at points where curved sections begin and where tubes end. So, an adaptive step size numerical solver will unnecessarily attempt to refine the mesh around these points. Thus, for reasons of numerical efficiency and repeatable accuracy, we recommend the process

of integrating each continuous section one at a time in series while algebraically enforcing continuity across the boundary from one section to the next.

Once one has solved (29) numerically to obtain all $\theta_i(s)$ and $\dot{\theta}_i(s)$, the result can be used to obtain $\boldsymbol{\omega}$ algebraically from (18). The resulting backbone trajectory of the cannula can be calculated by integrating $\dot{R}(s) = R^T(s)\hat{\boldsymbol{\omega}}(s)$ to get $R(s)$ and then integrating (2) to get $\mathbf{t}(s)$. There exist a number of numerical integration methods that preserve the structure of $SE(3)$, which can be used to obtain $R(s)$ from $\boldsymbol{\omega}(s)$. For a review of these techniques we refer the reader to (41).

In the remainder of this paper, we provide an analytical solution for an important special case of (23), where the active cannula consists of two circularly precurved tubes. A version of this analytical solution was first presented in (25), and subsequently discussed further in (26). This special case is interesting because an analytical solution exists, and because despite the fact that many-tube active cannula prototypes have been built, all experimental inquiries to date into active cannula behavior have addressed exclusively this particular case (see e.g. (1; 26)). Our general model (23) and more specifically the analytical solution of the special case, provides the first means for exploring the effects of torsion on active cannula shape.

4.2 Analytical Solution for Two Circular Tubes with Torsion

While (23) can be solved numerically for any number of component tubes, much insight into the fundamental behavior of active cannulas can be gained by considering the case of an active cannula composed of two circularly precurved tubes, for which an analytical solution can be found. Without loss of generality, one can define the two tube frames such that $\boldsymbol{\omega}_1^* = [\kappa_1 \ 0 \ 0]^T$ and $\boldsymbol{\omega}_2^* = [\kappa_2 \ 0 \ 0]^T$. Expanding (23) in this case yields

$$\frac{G_1 J_1 G_2 J_2}{G_1 J_1 + G_2 J_2} \ddot{\theta} - \kappa_1 \kappa_2 \frac{E_1 I_1 E_2 I_2}{E_1 I_1 + E_2 I_2} \sin \theta = 0, \quad (30)$$

where θ_2 has been replaced with θ . Under the assumption that the two tubes have the same value of ν , Poisson's ratio, this equation reduces to

$$\ddot{\theta} - \kappa_1 \kappa_2 (1 + \nu) \sin \theta = 0, \quad (31)$$

however, we will not make this assumption here, for the sake of generality. Let

$$a = \kappa_1 \kappa_2 \frac{E_1 I_1 E_2 I_2 (G_1 J_1 + G_2 J_2)}{G_1 J_1 G_2 J_2 (E_1 I_1 + E_2 I_2)}, \quad (32)$$

so that

$$\ddot{\theta} - a \sin \theta = 0. \quad (33)$$

The appropriate boundary conditions here are the initial angle determined by the relative angular position of the tube bases, $\theta(0) = \theta_0$, and the natural boundary condition at the free end, $\dot{\theta}(L) = 0$. This natural boundary condition can be intuitively understood by considering that $\dot{\theta}(s) \propto$ axial torque applied at s . At L there is no torque being applied to the distal end of either tube. Thus, $\dot{\theta}(L) = 0$.

Notice that (33) has the same form as the differential equation which describes a simple pendulum. Fortunately, this equation arises often, and it has a known analytical solution in terms of Jacobi's elliptic functions. We solve it following the solution procedure similar to the method described in (42), which begins by multiplying both sides of (33) by $\dot{\theta}$, and integrating once. Applying the boundary condition $\dot{\theta}_L = 0$ then yields

$$\dot{\theta}^2 + 2a \cos \theta = 2a \cos \theta_L. \quad (34)$$

Rearranging (34), we obtain

$$s = \pm \frac{1}{\sqrt{2a}} \int_{\theta_0}^{\theta(s)} \frac{d\theta}{\sqrt{\cos \theta_L - \cos \theta}} \quad (35)$$

Now we use $\cos(\theta + \pi(1 + 2n)) = -\cos \theta \quad \forall n \in \mathbb{Z}$, and $\cos 2\theta = 1 - 2\sin^2 \theta$, to write this in terms of incomplete elliptic integrals of the first kind as follows. First let

$$\gamma = \theta + \pi(1 + 2n), \quad k = \sin \frac{\gamma_L}{2} = \sqrt{\frac{1 - \cos \gamma_L}{2}}, \quad (36)$$

and let Φ be defined by

$$\cos \gamma = 1 - 2k^2 \sin^2 \Phi. \quad (37)$$

Then we have

$$\cos \gamma - \cos \gamma_L = 2k^2 \cos^2 \Phi, \quad \text{and} \quad \sin \gamma = 2k \sin \Phi (1 - k^2 \sin^2 \Phi)^{\frac{1}{2}} \quad (38)$$

So that (35) becomes

$$s = \pm \sqrt{\frac{1}{a}} \int_{\Phi(0)}^{\Phi(s)} \frac{d\theta}{\sqrt{1 - k^2 \sin^2(\Phi)}} = \pm \sqrt{\frac{1}{a}} \left(F(\Phi(s), k) - F(\Phi(0), k) \right) \quad (39)$$

where $F(\Phi, k)$ is the elliptic integral of the first kind with amplitude Φ and modulus k . We can now use the Jacobi Amplitude functions, sn and cn which are the sine and cosine of the inverse function of F defined by the identities

$$\text{sn}(F(\Phi, k), k) = \sin \Phi \quad \text{and} \quad \text{cn}(F(\Phi, k), k) = \cos \Phi \quad (40)$$

to obtain the following solutions:

$$\begin{aligned}\theta(s) &= 2 \sin^{-1} (k \operatorname{sn} (F(\Phi(0), k) \pm \sqrt{a}s, k)) - \pi(1 + 2n) \\ \dot{\theta}(s) &= \pm 2k\sqrt{a} \operatorname{cn} (F(\Phi(0), k) \pm \sqrt{a}s, k).\end{aligned}\tag{41}$$

The \pm signs in (41) and take the same sign as $\theta_L - \theta_0$, and n is chosen such that γ_L takes on a value in the range $-\pi \leq \gamma_L \leq \pi$.

4.3 Examples

In order to illustrate how to use the solution described above to obtain the shape of an active cannula, and to investigate the phenomena of multiple solutions and non-circular equilibrium shapes, we provide the following example. Consider a tube and a wire with properties given in Table 4.3. The long curved lengths of $L_1=200$ mm and $L_2=140$ mm for the inner wire and the outer tube, respectively, make the interaction of the curved portions pronounced and thus more easily visualizable. The boundary condition at the proximal end, where $s = 0$, is $\theta(0) = \theta_0 = \alpha_2 - \alpha_1$, where α_2 and α_1 are the base rotation angles applied by the actuators at the tube bases. The boundary condition at the free distal end where the outer tube ends ($s = L_2$) is $\dot{\theta}_{L_2} = 0$, which was already implicitly enforced on our solution in Equation (34). The analytical solution (41) contains the unknown constant, θ_{L_2} within k , so we need to find a solution for θ_{L_2} which satisfies the proximal boundary condition $\theta_0 = \alpha_2 - \alpha_1$. We can then think of “starting at the distal end” and use the change of variables $\sigma = L_2 - s$ to rewrite (41) in terms of θ_{L_2} as

$$\theta(\sigma) = 2 \sin^{-1} (k \operatorname{sn} (F(\Phi(L_2), k) \pm \sqrt{a}(\sigma), k)) - \pi(1 + 2n)\tag{42}$$

So that θ_{L_2} must be determined to satisfy

$$\alpha_2 - \alpha_1 - 2 \sin^{-1} (k \operatorname{sn} (F(\Phi(L_2), k) \pm \sqrt{a}(L_2), k)) - \pi(1 + 2n) = 0.\tag{43}$$

This can be done using standard nonlinear root finding techniques such as MatLab's `fzero`. Once θ_L is found, the analytical solution (41) is used to obtain $\theta(s)$, which enables us to obtain each tube's deformed curvature from (18) and (7). Then, the resulting shape of each tube can be obtained by numerically integrating $\dot{R}_i(s) = R_i(s)e^{\hat{\omega}(s)_i}$ and using (2) to determine $\mathbf{t}(s)$.

In (1) torsion was considered in straight transmission sections of a cannula. It was shown that in this case, multiple solutions (local minimum energy configurations) can emerge. We see the same phenomenon here. In general, there can be more than one value of θ_L which satisfies (43). The particular root which the algorithm converges to is dependent on the initial guess. As noted in (1) the particular configuration (solution) taken by the cannula will depend on actuator history. We illustrate this phenomenon in our example below. The inner wire is rotated to an angle of $\alpha_1 = -180^\circ$, while the outer tube stays at $\alpha_2 = 0^\circ$, making $\theta_0 = 180^\circ$. The left hand side of Equation 43 is depicted in Fig. 4 with respect to θ_{L_2} . Note that there are 3 places where the graph crosses the x -axis, representing three different solutions to (43), and corresponding to the three different configurations shown in Fig. 3. The solution at $\theta_{L_2} = 180^\circ$ is a trivial solution to the differential equation, representing the case where neither tube undergoes any torsion (a torsionally rigid model would produce this result). This is an unstable configuration in that the cannula will snap to one of the other solutions if perturbed slightly. The cannula will reach the $\theta_{L_2} = 84.4^\circ$ solution if the actuator starts at $\alpha_1 = 0^\circ$ and increases α_1 continuously until $\alpha_1 = 180^\circ$. If

	Outer Tube	Inner Wire
Young's Modulus (GPa)	58	58
Shear Modulus (GPa)	21.5	21.5
Inner Diameter (mm)	2.01	0
Outer Diameter (mm)	2.39	1.60
Length (mm)	140	200
Curvature (1/mm)	0.0099	0.0138

Table 1: Physical Properties of Tube and Wire used in Simulation.

the actuator decreases continuously from $\alpha_1 = 0^\circ$ to $\alpha_1 = -180^\circ$, the solution $\theta_{L_2} = -84.4^\circ$ will be reached, which corresponds to the $\theta_{L_2} = 275.6^\circ$ solution for $\alpha_1 = 180^\circ$. In order to solve (43) for the value of θ_{L_2} which corresponds to the actual configuration of the cannula, it is helpful to start simulating at a known configuration for which there is only one solution (e.g. $\theta_0 = 0$). Then, by undergoing incremental changes in θ_0 , the solution for θ_L at the previous step can be used as the initial guess for the current configuration. This results in the simulation portraying the same solution as the physical cannula until a bifurcation in the cannula energy is reached (where the current solution vanishes – see (1)) and the cannula




Figure 3: Simulation of the tubes given in Table 4.3 with the inner wire rotated to a base angle of $\alpha_1 = -180^\circ$ so that $\theta_0 = 180^\circ$. Three equilibrium conformations are shown corresponding to the three boundary condition solutions shown in Fig. 4. The solution with $\theta_{L_2} = 84.4^\circ$ is reached by rotating α_1 in the negative direction to $\alpha_1 = -180^\circ$, and the solution $\theta_{L_2} = 275.6^\circ$ is reached by rotating α_1 in the positive direction to $\alpha_1 = 180^\circ$. The solution with $\theta_L = 180^\circ$ is the trivial (unstable) solution, with the tubes undergoing no torsion.




Figure 4: Shown is a plot of the left hand side of Equation 43 versus θ_L for the tubes in Table 4.3 and $\theta_0 = 180^\circ$. Solutions for θ_L satisfying (43) are shown at $\theta_L = 180^\circ$, $\theta_L = 84.4^\circ$, and $\theta_L = 275.6^\circ$.

“snaps around” to a new solution.

For the a simple, two tube, circular precurvature case we are currently considering, it is possible to predict analytically when multiple solutions will exist (see (26) for an alternate derivation of the following result). For $\theta_0 = 180^\circ$ (the angular input where multiple solutions will first exist) we can examine the integral in Equation 39 to determine whether multiple solutions are possible. If $\theta_L = 180^\circ$, the integral is zero by definition, which means cannulas of any overlapped length have a solution $\theta_L = 180^\circ$. For $\theta_L \neq 180^\circ$, the integral has a lower bound of $\pi/2$, which can be seen in Fig. 5 and is shown in (44). Thus, by rearranging (39) and applying this inequality, it can be seen that for cannulas with a value of $L\sqrt{a} < \pi/2$, only the trivial solution exists. On the other hand, if $L\sqrt{a} \geq \pi/2$, two nontrivial solutions also exist, symmetric about $\theta_L = 180^\circ$. Thus, as shown in (45), the dimensionless parameter $L\sqrt{a}$ - which is composed of the overlapped length, stiffness, and curvature of the tubes -




Figure 5: For $\theta_0 = 180^\circ$ the value of the integral in (44) is shown in blue as a function of θ_L ranging from 0° to 360° . Because it is lower bounded by $\frac{\pi}{2}$, the dimensionless parameter $L\sqrt{a}$ can be used to predict when multiple solutions can occur.

can be used to predict whether a two tube cannula will exhibit multiple solutions and thus have the potential to “snap” from one stable solution to another.

To summarize our multiple solutions discussion above, we have the inequality

$$\pi/2 \leq \left| \int_0^{\Phi(L)} \frac{d\theta}{\sqrt{1 - k^2 \sin^2(\Phi)}} \right| \quad \forall \theta_L \neq 180^\circ, \quad (44)$$

which when combined with (39) yields the conditions:

$$\begin{aligned} L\sqrt{a} < \pi/2 &\rightarrow \text{only one solution} \\ L\sqrt{a} \geq \pi/2 &\rightarrow \text{multiple solutions exist} \end{aligned} \quad (45)$$

As shown in Figure 3, the two solutions with torsion are significantly different than the




Figure 6: Shown above are four configurations of a simulation of two fully precurved, fully overlapping tubes, whose material properties are given in Table 4.3. Both tubes have a longer arc length of 636.5 mm(equal to one full circle of the outer tube). The inner wire is rotated in the positive direction to angles of 90° , 225° , 315° , and 350° at the base. It is evident that in extreme cases, circular tubes with precurvature can form highly noncircular shapes when combined due to the effects of torsion.

no torsion solution. Thus torsion in the overlapping curved section can be very important for determining overall shape. Still, it would appear that each section is very close to circular, suggesting that some kind of adjustment to a piecewise circular model could be an effective way to compensate for torsion in the curved sections. Modeling the individual sections of an active cannula as circular arcs leads to very convenient kinematic formulations that have been widely exploited in prior work (e.g.(1; 10; 13), etc. – see also (43) for an overview of piecewise constant curvature kinematics for continuum robots). However, the presence of torsion can, in some cases, lead to curved shapes that are qualitatively different and

which cannot be approximated well by circular arcs. If the overlapped arc length is long or the curvatures are large, torsional relaxation makes it possible to obtain highly non-circular shapes from two circularly precurved tubes. To illustrate this, we extend the curved portions of both tubes used in our first example to 636.5 mm (corresponding to one full circle of the outer tube) and rotate the inner wire from $\alpha_1 = 0^\circ$ to $\alpha_1 = 350^\circ$. The resulting shape is shown in Figure 6, where the inner wire has been rotated in the positive direction to angles of 90° , 225° , 315° , and 350° at the base. It is clear from Figure 6 that the resulting shape cannot be well approximated by a circle. However, for cannulas of sufficiently short curved overlap and sufficiently small curvature, piecewise circular models are reasonably accurate at predicting cannula shape. Such was the case for many prior prototypes (e.g. (1; 13)).

5 Experimental Validation of the Analytical Model

The range of possible diameters for active cannula robots is defined by the diameters of elastic tubes that can be manufactured, shaped, and actuated. The prototype used in this study is comprised of superelastic Nitinol tubes, but any material with a high recoverable strain that can be appropriately shaped (e.g. plastics) may be used. Nitinol tubes can be shape-set by heat treatment or plastic deformation, and are available in stock sizes ranging from 0.2 mm




Figure 7: Diagram of tube overlap configuration with variables from (46) shown.

to 8.0 mm (44). Furthermore, the dexterity of the design improves with miniaturization because smaller tubes can sustain higher precurvatures (12).

In this section we will compare the predictions of three available models for active cannula shape with a set of experiments on a prototype cannula. The three models are the “bending only” model (13), the “transmissional torsion” model (12; 14; 1), and the model provided in Section 4, a preliminary version of which was presented in (27). The experimental data set used here is the same as that provided in (1).

5.1 Model Implementation

As described in (1), these experiments use an active cannula constructed of one tube and one wire, each of which has an initial straight transmission, followed by a circularly curved section near its tip. Thus, as shown in Figure 7, this prototype can be considered to have four distinct regions or “links”, which begin and end where tubes begin and end, or where tubes transition from straight to curved. For example, beginning at the base of the cannula, the links will often be as follows: (1) a link where both tubes are straight, (2) a link where one tube is curved and the other is straight, (3) a link where both tubes are curved, and (4) a link where only one tube is present and curved.

Since our experimental prototype contains straight transmission segments that connect the actuators to the start of the curved tube sections, our model implementation must account for the torsional compliance of the transmission in addition to that of the curved sections. In order to use the general model directly as derived in Section 2, we will (without loss of generality) consider $s = 0$ to refer to the point where both tubes become curved (the beginning of link 2 – labeled ℓ_2 – in Figure 7). We need to modify the boundary condition given by (43) to account for the straight length of each tube before it becomes curved. We know that the angle of twist for both tubes varies linearly with arc length in the sections which have at least one straight tube (the trivial solution of (6) with $n = 2$). Using this and the fact that the moments about the z axes to achieve α_1 and α_2 must balance, we can

obtain

$$\dot{\theta}_0 - (\theta_0 - \alpha_1 + \alpha_2) \frac{G_1 J_1 + G_2 J_2}{G_1 J_1 D_2 + G_2 J_2 D_1} = 0, \quad (46)$$

thus (46) is to be satisfied instead of (43) where θ_0 and $\dot{\theta}_0$ are expressed as functions of the unknown value θ_{L_2} as follows

$$\begin{aligned} \theta_0(\theta_{L_2}) &= 2 \sin^{-1} (k \operatorname{sn} (F(\Phi(L_2), k) \pm \sqrt{a}(L_2), k)) - \pi(1 + 2n) \\ \dot{\theta}_0(\theta_{L_2}) &= \pm 2k\sqrt{a} \operatorname{cn} (F(\Phi(L_2), k) \pm \sqrt{a}(L_2), k), \end{aligned} \quad (47)$$

where D_1 and D_2 are the arc lengths between the actuators and the link in which both tubes are curved, and α_1 and α_2 are the tube base input angles as shown in Figure 7.

5.2 Experimental Dataset

Here, we summarize the experimental data set (the dataset from (1)) which was used to compare the model given in Section 2 with transmisional torsion model. In these experiments, an outer tube and an inner wire were arranged in two different translational positions and a range of input angles were applied. The two translational positions were referred to as the “full overlap case” and the “partial overlap case”. In the full overlap case, the tube and wire were arranged so that the link lengths were as follows: $\ell_1 = 10\text{mm}$ (tube curved, wire straight), $\ell_2 = 82.3\text{mm}$ (both curved), and $\ell_3 = 2.7\text{mm}$ (only wire present). For the full overlap case, data was recorded for 15 different input angles ranging from 0° to 280° in 20° increments. In the partial overlap case, the tube and wire were arranged so that the link lengths were as follows: $\ell_1 = 48\text{mm}$, $\ell_2 = 44.3\text{mm}$, and $\ell_3 = 40.7\text{mm}$. For this overlap configuration, data was recorded for 11 different input angles ranging from 0° to 200° in 20° increments. The workspace range covered by these actuation inputs is depicted in Figure 9.




Figure 8: Manual actuation mechanism used in experiments. In this apparatus, both tube and wire are affixed to circular acrylic input handles at their bases, which are etched to encode rotation. The support structure is etched with a linear ruler to encode translation. Spring pin locking mechanisms lock the input discs at desired linear and angular input positions. The inset image of a striped cannula on a white background is an example of an image captured using one of our calibrated stereo cameras. The black bands seen are electrical tape and allow for point correspondences to be identified for stereo triangulation. The red circles indicate the locations at which euclidean errors were calculated. Calibration of model parameters was done to minimize the sum of these errors over all experiments.

5.3 Procedure and Model Calibration

Each of the above configurations were input to the base of the cannula tube and wire using the manual actuation unit shown in Figure 8. For each, the resultant overall shape of the cannula was recorded via a calibrated pair of stereo cameras (Sony XCD-X710 firewire cameras with a resolution of 1024×768 pixels). The fiducial markers shown in the inset

	Outer Tube	Inner Wire
Young's Modulus (GPa)	58	58
Shear Modulus (GPa)	21.5	21.5
Inner Diameter (mm)	2.01	0
Outer Diameter (mm)	2.39	1.60
Straight Length (mm)	93.5	218.5
Curved Length (mm)	92.3	85
Curvature (1/mm)	0.0099	0.0138

Table 2: Measured and Assumed Physical Quantities for Experimental Tube and Wire.

image in Figure 8 enabled determination of point correspondences for stereo triangulation, after they had been identified in image coordinates by manually clicking on the center of the black bands in video frames. One source of error in this data collection procedure is the accuracy of manual point selection in images, which is estimated to be approximately 2 pixels or 0.6 mm. Another is fiducial size (they are not infinitesimal points), causing small differences in intended selection locations. We estimate that fiducial dimensions introduce error of approximately the diameter of the wire itself (1.6 mm). Based on these, our overall vision system measurement error is approximately 2.2mm, in a worst-case sense.

The nominal physical properties of the tube and wire used in our experiments are given in Table 5.2. We compare the model of Section 4 with the prior transmisional torsion model using both the nominal values given in Table 5.2 and calibrated parameters (a calibration procedure for the transmisional torsion model is provided in (1)). Examining the equations in Section 4, we see that the stiffness coefficients in Equations 18 and 33 can be expressed in terms of the three dimensionless parameters

$$c_1 = \nu_1, \quad c_2 = \nu_2, \quad \text{and} \quad c_3 = \frac{E_1 I_1}{E_1 I_1 + E_2 I_2}. \quad (48)$$

Poisson's ratio is often taken to be approximately 0.35 for Nitinol. It has also been




Figure 9: Configuration space covered in experiments. Left: partial overlap case, Right: full overlap case.

noted that plastic deformation can increase Poissons ratio for Nitinol to 0.5 (45). Since we pre-shaped our tubes via plastic deformation, we will assume a range of 0.30 to 0.50. An expected range for c_3 can be deduced from the uncertainty in each quantity upon which it depends. Nitinol dimensions are specified by the manufacturer (Nitinol Devices and Components, Inc.) to ± 0.0010 in., while the elastic modulus E is reported as 41 to 75 GPa. Applying standard error propagation, the expected range for parameter c_3 is 0.143 to 0.431. These ranges provide a basis for comparison with fitted parameter values produced by the parameter fitting procedure. To calibrate the parameters, we minimize the sum of the positional errors at the tip of the wire, the tip of the tube and the measured point most near the base, as shown in Figure 8. Matlab's `fmincon` function was used to optimize the values of the three dimensionless parameters given in (48) with upper and lower bounds set to the expected ranges of the parameters. As described in (1), the transformation between the stereo camera coordinate frame and a frame fixed at the base of the cannula was first estimated using point cloud registration (46). Images of a 15-mm checkerboard pattern (with corners at known physical locations with respect to the cannula base frame) were captured. Sixteen corners on the checkerboard were triangulated with the stereo vision system. This registration was only expected to provide a rough estimate of the frame transformation. Thus, six nuisance parameters (a 3-vector for position and a 3-vector for orientation with magnitude of rotation encoded as length) describing the cannula base frame were included in the calibration procedure and initialized with the results from the point cloud registration. Nuisance parameters showed only small changes during optimization, with cannula base frame moving only 0.5 mm, and rotating through X-Y-Z Euler Angles of $\alpha = 0.9^\circ$, $\beta = 0.3^\circ$, $\gamma = 4.0^\circ$.

Parameter	Nominal Value	Calibrated Value
c_1	0.350	0.451
c_2	0.350	0.449
c_3	0.287	0.341

Table 3: Nominal and Calibrated Dimensionless Parameters




Figure 10: Comparison of shape for the transmisional torsion model (green – dotted line) with nominal parameters, the model given in Section 4 (red – solid line) with nominal parameters, and experimental data (blue – dashed line) for configurations near the edge of the active cannula workspace. Note that the model given in Section 4 produces predictions closer to experimentally observed cannula shape. Left: partial overlap case, Right: full overlap case.

5.4 Results

The calibrated parameter values are given alongside their nominal values in Table 5.3, and we note that they fall well within their expected ranges and converge to near the same values for initial guesses in a range within $\pm 5\%$ of the optimal values. In (1) calibration led to one of the parameters falling outside its expected range, which illustrates that the model presented in this paper more completely captures the underlying mechanics. Note that the unmodeled presence of friction would have a similar effect on our data as lowering the torsional rigidity of the tubes, namely increasing torsional windup. Thus, the calibration process would tend

	Mean Tip Error (mm)	Max Tip Error (mm)
Bending Only Model	24.8	54.3
Transmisional Torsion Model	10.1	22.1
Model of Section 4	4.7	12.7

Table 4: Uncalibrated Tip Error Statistics for Current Model Compared to Prior Models




Figure 11: Comparison of shape for the transmisional torsion model (green – dotted line) with calibrated parameters, the model given in Section 4 (red – solid line) with calibrated parameters, and experimental data (blue – dashed line) for configurations near the edge of the active cannula workspace. Note that the model given in Section 4 produces predictions closer to experimentally observed cannula shape. Left: partial overlap case, Right: full overlap case.

to increase c_1 and c_2 to compensate for frictional effects. This may account for the slightly high values of c_1 and c_2 , but they are nevertheless still within their expected ranges.

This is also supported by the data in Table 5.4. When using nominal parameters from data sheets, the model of Section 4 is significantly more accurate than the transmisional torsion model. Specifically, the model of Section 4 results in an average tip error of only 4.72 mm as opposed to 10.1 mm for the transmisional model. Figure 10 shows the experimental data and the predictions of both models using nominal parameters for the two worst experimental cases, where angular input angle differences are at the edge of the workspace and

	Mean Tip Error (mm)	Max Tip Error (mm)
Bending Only Model	13.6	31.5
Transmisional Torsion Model	3.0	8.8
Model of Section 4	2.5	7.1

Table 5: Calibrated Tip Error Statistics for Current Model Compared to Prior Models

torsion is most significant. These are 280° in the full overlap case, and 200° in the partial overlap case.

Quantitatively, the model of Section 4 with calibrated parameters exhibits a mean tip error of 2.5 mm across all experiments with a maximum tip error of 7.1 mm, as shown in Table 5.4). In comparison, the calibrated transmissional torsion model exhibits a mean of 3.0 mm and a maximum of 8.8 mm, and the bending only model a mean of 13.6 mm and a maximum of 31.5 mm. With calibrated parameters, the model of Section 4 improves the mean tip error 82% over the bending only model, and 17% over the transmissional torsion model.

Plots of the experimental data and the predictions of both models using calibrated parameters are shown in Figure 11, picturing the same two “worst-case” experiments shown in Figure 10. The behavior pictured is common to all experimental positions using either calibrated or uncalibrated parameters, namely that the prediction of the model of Section 4 lies nearer the experimental data than the transmissional torsion model prediction. Note also that the predictions of the bending only model are not shown in for clarity, since they are sufficiently far from the experimental data that they would obscure the differences between the other two models. As discussed in (1), the bending only model neglects the torsional windup that occurs in an active cannula, so its predictions become increasingly structurally incorrect as the angle input difference increases.

To demonstrate the fact that tip error is a good metric to use, we give the error between




Figure 12: Error vs. arc length for the 200° partial overlap case with nominal parameter values using the model of Section 4. The general increase in error from base to tip is characteristic of all experiments, and thus tip error provides a reasonable metric for our experimental dataset.




Figure 13: Angle ϕ of that defines the resulting instantaneous plane of curvature of the active cannula. Left: partial overlap case, Right: full overlap case.

the model prediction with nominal parameters and the experimental backbone location in Figure 12 as a function of arc length for the worst experimental case (the left case in Figure 11). It shows the common characteristic that the positional error increases with arc length.

Using the model of Section 4, Figure 13 shows the angle defining the plane of curvature of the experimental cannula as a function of arc length. The plot shows that while this angle is not exactly piecewise constant, it is approximately so. This illustrates why previous studies (e.g. (12; 13)) have successfully used models which imply piecewise constant curvature. The out of plane motion was small enough that it was not obvious experimentally, and cannula shape was approximately piecewise circular. These quantitative results also indicate that the general modeling framework is providing enhanced predictive ability by reducing tip error, and thus may expand the range of medical procedures to which active cannulas can be applied.

6 Conclusions

We have presented a general coordinate-free energy formulation for modeling the shape of concentric tube continuum robots known as active cannulas. This formulation is able to account for precurvatures and stiffnesses that vary along the length of component tubes,

and to explicitly model torsion throughout the device. Further, previous models are special cases of this new general modeling framework, and we showed that the moment balance equations assumed in prior work arise naturally within our new modeling framework. In this work we also showed via simulation and experiments that the new model described in this paper captures the underlying mechanics of the cannula more accurately than prior models. This was illustrated in simulation by the fact that tubes with circular precurvatures can combine to form a dramatically different shape if torsion is allowed in curved sections. In the experiments it was shown that the calibrated parameter values fell within their expected ranges, which was not the case for the model with only transmisional torsion.

This new model may have significant implications for active cannula applications in both medicine and industry. In this paper we have not addressed design guidelines for active cannulas, a discussion of which can be found in (1; 16). Rather, we have provided the model necessary to simulate possible designs and judge their merits in comparison to design goals. Explicitly accounting for torsion throughout the cannula can significantly enhance the model, enabling active cannulas to be used in more demanding applications. Furthermore, active cannulas with variable precurvatures will be able to reach further and through more complex trajectories while using fewer tubes. This new model also facilitates future studies on patient-specific preshaping of active cannula component tubes, so that one may match the capabilities of the device to the particular location and entry trajectory required by a specific patient. An important area for future work is the development of active cannula models that consider external loading. This may be accomplished through use of the Euler–Poincaré approach, or by application of Cosserat rod theory. As suggested in (47) and partially explored in (48), such a future model model will be useful for determining cannula shape when it is interacting with tissue, using the cannula itself as a force sensor (by sensing shape), and implementing force controllers. If it becomes necessary in future models to take into account the effects of shear and elongation or other complicating factors, the Euler–Poincaré approach may be particularly valuable.

Acknowledgement

This work was supported in part by the National Science Foundation under #CBET-0651803.

References

- [1] R. J. Webster III, J. M. Romano, and N. J. Cowan, “Mechanics of precurved-tube continuum robots,” *IEEE Transactions on Robotics*, vol. 25, no. 1, pp. 67–78, 2009.
- [2] M. W. Hannan and I. D. Walker, “Kinematics and the implementation of an elephant’s trunk manipulator and other continuum style robots,” *Journal of Robotic Systems*, vol. 20, no. 2, pp. 45–63, 2003.
- [3] P. Dario, M. C. Carrozza, M. Marcacci, S. D’Attanasio, B. Magnami, and G. M. O. Tonet, “A novel mechatronic tool for computer-assisted arthroscopy,” *IEEE Transactions on Information Technology in Biomedicine*, vol. 4, no. 1, pp. 15–29, 2000.
- [4] D. B. Camarillo, C. F. Milne, C. R. Carlson, M. R. Zinn, and J. K. Salisbury, “Mechanics modeling of tendon-driven continuum manipulators,” *IEEE Transactions on Robotics*, vol. 24, no. 6, pp. 1262–1273, 2008.
- [5] N. Simaan, R. Taylor, and P. Flint, “A dexterous system for laryngeal surgery,” *IEEE International Conference on Robotics and Automation*, pp. 351–357, 2004.
- [6] G. S. Chirikjian, “Kinematic synthesis of mechanisms and robotic manipulators with binary actuators,” *Journal of Mechanical Design*, vol. 117, pp. 573–580, 1995.
- [7] B. A. Jones, W. McMahan, and I. D. Walker, “Practical kinematics for real-time implementation of continuum robots,” *IEEE International Conference on Robotics and Automation*, pp. 1840–1847, 2006.

[8] M. Loser, “A new robotic system for visually controlled percutaneous interventions under X-ray or CT-fluoroscopy,” Master’s thesis, The Albert-Ludwig-University, Freiburg, Germany, September 2002.

[9] W. Daum, “A deflectable needle assembly,” 2003, patent 6,572,593.

[10] J. Furusho, T. Katsuragi, T. Kikuchi, T. Suzuki, H. Tanaka, Y. Chiba, and H. Horio, “Curved multi-tube systems for fetal blood sampling and treatments of organs like brain and breast,” *Journal of Computer Assisted Radiology and Surgery*, pp. 223–226, 2006.

[11] M. Terayama, J. Furusho, and M. Monden, “Curved multi-tube device for path-error correction in a needle-insertion system,” *International Journal of Medical Robotics and Computer Assisted Surgery*, vol. 3, no. 2, pp. 125–134, 2007.

[12] R. J. Webster III, A. M. Okamura, and N. J. Cowan, “Toward active cannulas: Miniature snake-like surgical robots,” *IEEE/RSJ International Conference on Intelligent Robots and Systems*, pp. 2857–2863, October 9-15 2006.

[13] P. Sears and P. E. Dupont, “A steerable needle technology using curved concentric tubes,” *IEEE/RSJ International Conference on Intelligent Robots and Systems*, pp. 2850–2856, October 9-15 2006.

[14] R. J. Webster III, J. M. Romano, and N. J. Cowan, “Kinematics and calibration of active cannulas,” *IEEE International Conference on Robotics and Automation*, pp. 3888–3895, 2008.

[15] P. Sears and P. E. Dupont, “Inverse kinematics of concentric tube steerable needles,” *IEEE International Conference on Robotics and Automation*, pp. 1887–1892, 2007.

[16] R. J. Webster III, “Design and mechanics of continuum robots for surgery,” Mechanical Engineering, Johns Hopkins University, Baltimore, MD, December 2007, PhD Thesis.

[17] K. Ediz and N. Olgac, “Microdynamics of the piezo-driven pipettes in icsi,” *IEEE Transactions on Biomedical Engineering*, vol. 51, no. 7, pp. 1262 – 1268, 2004.

[18] Y. Sun and B. J. Nelson, “Biological cell injection using an autonomous microrobotic system,” *The International Journal of Robotics Research*, vol. 21, no. 10-11, pp. 861–868, 2002.

[19] R. Kumar, A. Kapoor, and R. H. Taylor, “Preliminary experiments in robot/human cooperative microinjection,” *IEEE/RSJ International Conference on Intelligent Robots and Systems*, pp. 3186– 3191, 2003.

[20] S. P. DiMaio and S. E. Salcudean, “Interactive simulation of needle insertion models,” *IEEE Transactions on Biomedical Engineering*, vol. 52, no. 7, pp. 1167–1179, July 2005.

[21] D. Glozman and M. Shoham, “Image-guided robotic flexible needle steering,” *IEEE Transactions on Robotics*, vol. 23, no. 3, pp. 459–467, June 2007.

[22] R. Alterovitz, K. Goldberg, J. Pouliot, R. Taschereau, and I.-C. Hsu, “Sensorless planning for medical needle insertion procedures,” in *IEEE/RSJ International Conference on Intelligent Robots and Systems*, 2003, pp. 3337–3343.

[23] R. J. Webster III, J. S. Kim, N. J. Cowan, G. S. Chirikjian, and A. M. Okamura, “Non-holonomic modeling of needle steering,” *International Journal of Robotics Research*, vol. 25, no. 5/6, pp. 509–526, May/June 2006.

[24] S. Okazawa, R. Ebrahimi, J. Chuang, S. E. Salcudean, and R. Rohling, “Hand-held steerable needle device,” *IEEE/ASME Transactions on Mechatronics*, vol. 10, no. 3, pp. 285 – 296, 2005.

[25] D. C. Rucker and R. J. Webster III, “Mechanics-based modeling of bending and torsion in active cannulas,” *IEEE RAS/EMBS International Conference on Biomedical Robotics and Biomechatronics*, pp. 704–709, 2008.

[26] P. E. Dupont, J. Lock, and E. Butler, “Torsional kinematic model for concentric tube robots,” *IEEE International Conference on Robotics and Automation*, pp. 3851 – 3858, 2009.

[27] D. C. Rucker and R. J. Webster III, “Mechanics of bending, torsion, and variable precurvature in multi-tube active cannulas,” *IEEE International Conference on Robotics and Automation*, pp. 2533 – 2537, 2009.

[28] G. S. Chirikjian and J. W. Burdick, “Kinematically optimal hyper-redundant manipulator configurations,” *IEEE Transactions on Robotics and Automation*, vol. 11, no. 6, pp. 794–806, December 1995.

[29] I. A. Gravagne and I. D. Walker, “Kinematic transformations for remotely-actuated planar continuum robots,” *IEEE International Conference on Robotics and Automation*, pp. 19–26, 2000.

[30] I. A. Gravagne, “Design, analysis and experimentation: the fundamentals of continuum robotic manipulators,” Ph.D. dissertation, Clemson University, 2002.

[31] J. Selig and X. Ding, “A screw theory of static beams,” *IEEE/RSJ International Conference on Intelligent Robots and Systems*, vol. 1, pp. 312 – 317, 2001.

[32] S. S. Antman, *Nonlinear problems of elasticity*, 2nd ed. New York: Springer, 2005.

[33] R. L. Bishop, “There is more than one way to frame a curve,” *The American Mathematical Monthly*, vol. 82, no. 3, pp. 246–251, 1975.

[34] G. Chirikjian and A. Kyatkin, *Engineering Applications of Noncommutative Harmonic Analysis*. Boca Raton, FL: CRC Press, 2001.

[35] S. Kehrbaum and J. H. Maddocks, “Elastic rods, rigid bodies, quaternions and the last quadrature,” *Philosophical Transactions: Mathematical, Physical and Engineering Sciences*, vol. 355, p. 21172136, 1997.

- [36] M. Bergou, M. Wardetzky, S. Robinson, B. Audoly, and E. Grinspun, “Discrete elastic rods,” *SIGGRAPH (ACM Transactions on Graphics)*, 2008.
- [37] J. S. Kim and G. S. Chirikjian, “Conformational analysis of stiff chiral polymers with end-constraints,” *Molecular Simulation*, vol. 32, no. 14, pp. 1139–1154, December 2006.
- [38] G. S. Chirikjian, *Stochastic Models, Information Theory, and Lie Groups*. Birkhauser, Boston, 2011, vol. 2.
- [39] R. M. Murray, Z. Li, and S. S. Sastry, *A Mathematical Introduction to Robotic Manipulation*. Boca Raton, FL: CRC Press, 1994.
- [40] D. C. Rucker and R. J. Webster III, “Parsimonious evaluation of concentric-tube continuum robot equilibrium conformation,” *IEEE Transactions on Biomedical Engineering*, vol. 56, pp. 2308 – 2311, 2009.
- [41] J. Park and W.-K. Chung, “Geometric integration on euclidean group with application to articulated multibody systems,” *IEEE Transactions on Robotics*, vol. 21, no. 5, pp. 850–863, 2005.
- [42] W. F. Ames, *Nonlinear Ordinary Equations in Transport Processes*, R. Bellman, Ed. Academic Press Inc., 1968.
- [43] R. J. Webster III and B. A. Jones, “Design and modeling of constant curvature continuum robots: A review,” *International Journal of Robotics Research*, 2010, (Accepted).
- [44] Nitinol Devices and Components, Inc., www.nitinol.com, stock Lists. Last accessed August 3, 2008.
- [45] P. M. White, “Stress induced interposed connector,” 2001, patent 6,257,593.
- [46] K. S. Arun, T. S. Huang, and S. D. Blostein, “Least-squares fitting of two 3-D point sets,” *IEEE Transactions on Pattern Analysis and Machine Intelligence*, vol. 9, no. 5, pp. 698–700, 1987.

- [47] M. Mahvash and P. Dupont, “Bilateral teleoperation of flexible surgical robots,” *Workshop Proceedings, New Vistas and Challenges in Telerobotics. IEEE International Conference on Robotics & Automation*, pp. 58–64, 2008.
- [48] D. C. Rucker, B. Jones, and R. J. Webster III, “A model for concentric tube continuum robots under applied wrenches,” *IEEE International Conference on Robotics and Automation*, 2010, in press.